

Table 3
Reduction Factors

Test Substance	Contact Time	Replicate	Initial Load (Log ₁₀ TCID ₅₀)*	Output Load (Log ₁₀ TCID ₅₀)	Log ₁₀ Reduction	Reduction (%)
Pretz-MD Saline Solution	5 minutes	Rep 1	6.83	4.21	2.63	99.76
		Rep 2	6.83	3.96	2.88	99.87

* The Virus Recovery Control Average was used as the Initial Load

CONCLUSIONS

When tested as described, Pretz-MD Saline Solution. Lot No. 0417X, was evaluated for its ability to inactivate Severe Acute Respiratory Syndrome-related Coronavirus (SARS CoV 2) (COVID-19 Virus). The results are presented in Tables 1-3.

All of the controls met the criteria for a valid test.